

Analytical Report Nr.

AR-20-YL-008064-02-EN

Sample code Nr.

560-2020-00008104

Date

25/11/2020

ANALYTICAL REPORT**Client Information**

Tejidos Rebes, S.L.
C7 Luxemburgo 47-53
Badalona SPAIN
+34619218574
andres@tejidosrebes.com

For the attention of Mr. Andrés Castellort

Sample Information

Order Code: EUAA70-00009039
Reception Date: 4-Nov-2020
Analysis Starting Date: 4-Nov-2020
Analysis Ending Date: 24-Nov-2020
Sample code Nr. 560-2020-00008104
Sample described as: Mask material/Material mascarilla

Requirements and decision rule

Customer requirements: UNE 0065 Mascarillas higiénicas reutilizables
Decision Rule: Shared risk - Simple acceptance.

Information provided by the customer*

Client Reference: MICROFIBRA 86 TECNOSANI BLANCO+ POPELIN VIENA TEC
Sample Description:
Purchase Order Number:

Batch No suministrado / Not provided

(*this report cancels and replaces the previous one, numbered AR-20-YL-008064-01/560-2020-00008104 dated 24/11/2020 which must be destroyed)

Analytical Report Nr.

AR-20-YL-008064-02-EN

Sample code Nr.

560-2020-00008104

Date

25/11/2020

SAMPLE PICTURE

Analytical Report Nr.

AR-20-YL-008064-02-EN

Sample code Nr.

560-2020-00008104

Date

25/11/2020

CONCLUSION:

TEST PROPERTY	PASS	FAIL	REMARKS
• BACTERIAL FILTRATION EFFICIENCY (BFE) EN 14683:2019+AC:2019 Annex B			
A - Original / As received	X		
B - Tras 50 lavados / After 50 washes	X		
◆ Washing - Reusable masks ISO 6330:2012			
B - Tras 50 lavados / After 50 washes			REFER RESULT
Breathability (Differential Pressure) EN 14683:2019+AC:2019 Annex C			
A - Original / As received	X		
B - Tras 50 lavados / After 50 washes	X		

Remark: Test has been performed as per application request

Analytical Report Nr.

AR-20-YL-008064-02-EN

Sample code Nr.

560-2020-00008104

Date

25/11/2020

COMPONENT LIST:

COMPONENT ID	COMPONENT NAME	MATERIAL DESCRIPTION	COLOR	REMARKS
CUST 01	A - Original / As received	Textil / Fabric	Blanco / White	---
CUST 02	B - Tras 50 lavados / After 50 washes	Textil / Fabric	Blanco / White	---

Analytical Report Nr.

AR-20-YL-008064-02-EN

Sample code Nr.

560-2020-00008104

Date

25/11/2020

MASKS TESTING	CAS No.	RESULTS	UNC.	LOQ	GUIDELINES
---------------	---------	---------	------	-----	------------

Analyses on: A - Original / As received

Breathability (Differential Pressure)

Analysis Ending Date: 24/11/2020

EN 14683:2019+AC:2019 Annex C

Differential pressure

36.2 Pa/cm² (± 2.0) Pa/cm²

-

<60 Pa/cm²

✓ Pass

Complete test data reported at Annex.

• BACTERIAL FILTRATION EFFICIENCY (BFE)

Analysis Ending Date: 24/11/2020

EN 14683:2019+AC:2019 Annex B

Bacterial Filtration Efficiency
(BFE)

96.4

-

≥90 %

✓ Pass

Complete test data reported at Annex.

Testing covered by NABL accreditation n°TC-5492.

Analyses on: B - Tras 50 lavados / After 50 washes

Breathability (Differential Pressure)

Analysis Ending Date: 24/11/2020

EN 14683:2019+AC:2019 Annex C

Differential pressure

49.9 Pa/cm² (± 1.5) Pa/cm²

-

<60 Pa/cm²

✓ Pass

Complete test data reported at Annex.

◆ Washing - Reusable masks

Analysis Ending Date: 24/11/2020

ISO 6330:2012

Number of washing cycles

50

-

Type of washing machine

A1

-

Temperature

60 °C

-

Washing procedure

6N

-

Drying method

C - Flat dry

-

Detergent used

Non Phosphate ECE 98 +
TAED + Sodium Perborate

-

Specimen and ballast total dry
mass

2 kg

-

Type of ballast

Type III - 100% Polyester

-

Analytical Report Nr.

AR-20-YL-008064-02-EN

Sample code Nr.

560-2020-00008104

Date

25/11/2020

MASKS TESTING	CAS No.	RESULTS	UNC.	LOQ	GUIDELINES
• BACTERIAL FILTRATION EFFICIENCY (BFE)					Analysis Ending Date: 24/11/2020
EN 14683:2019+AC:2019 Annex B					
Bacterial Filtration Efficiency (BFE)		95.8		-	≥90 % ✓ Pass

Complete test data reported at Annex.
Testing covered by NABL accreditation n°TC-5492.

Analytical Report Nr.

AR-20-YL-008064-02-EN

Sample code Nr.

560-2020-00008104

Date

25/11/2020

Reason for new version: The cust of the test after washes is modified due to error at registration.

Analytical Report Nr.

AR-20-YL-008064-02-EN

Sample code Nr.

560-2020-00008104

Date

25/11/2020

Signed for and on behalf of Eurofins Textile Testing Spain:

Report electronically validated by

Axel Ferrando

Physical-Mechanical Lab Manager

EXPLANATORY NOTE

- ◆ Test not covered by ENAC accreditation scope
- Test is subcontracted within Eurofins group and is accredited
- Test is subcontracted within Eurofins group and is not accredited
- Test is subcontracted outside Eurofins group and is accredited
- Test is subcontracted outside Eurofins group and is not accredited

N/A = Not Applicable

*Eurofins Textile Testing Spain S.L.U is not responsible of the information supplied by the costumer and reported as section "Information provided by the costumer".

Eurofins General Sales Terms and Conditions Applied.

Results obtained refer only to samples, products or material received in Laboratory, as described in section "Sample information" and tested in conditions shown in present report.

Test uncertainties not reported are at customer disposal, for those tests in which it is possible to evaluate the test uncertainty.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, which for a normal distribution provides a level of confidence of approximately 95%.

Reproduction of this document is only valid if it is done completely and under the written permission. Any printed version of this document will be considered as a copy.

If you happen to have any comments, please do it by sending email to textile_spain@eurofins.com and referring to this report number.

End Of Report

Código de Informe de Ensayo: AR-20-YL-008064-02-ES
Código de muestra 560-2020-00008104
Fecha de emisión 25/11/2020

INFORME DE ENSAYO

Información del cliente

Tejidos Rebes, S.L.
C7 Luxemburgo 47-53
Badalona ESPAÑA
+34619218574
andres@tejidosrebes.com

A la atención de Mr. Andrés Castellort

Información de la muestra

Código de la orden: EUAA70-00009039
Fecha de recepción de la muestra: 4-Nov-2020
Fecha de inicio de análisis: 4-Nov-2020
Fecha fin de análisis: 24-Nov-2020
Código de muestra 560-2020-00008104
Descripción de la muestra: Mask material/Material mascarilla

Requisitos y regla de decisión

Requerimientos del cliente: UNE 0065 Mascarillas higiénicas reutilizables
Regla de decisión: Riesgo compartido - Aceptación simple.

Información suministrada por el cliente*

Referencia de cliente: MICROFIBRA 86 TECNOSANI BLANCO+ POPELIN VIENA TEC
Descripción de la muestra:
Guía de remisión:

Lote No suministrado / Not provided

(* Este informe cancela y reemplaza el informe anterior, con número AR-20-YL-008064-01/560-2020-00008104 y fecha de 24/11/2020 que debe ser destruido)

Código de Informe de Ensayo:	AR-20-YL-008064-02-ES
Código de muestra	560-2020-00008104
Fecha de emisión	25/11/2020

FOTOGRAFÍA DE LA MUESTRA



Código de Informe de Ensayo: AR-20-YL-008064-02-ES
Código de muestra: 560-2020-00008104
Fecha de emisión: 25/11/2020

CONCLUSIÓN:

ENSAYO	CUMPLE	NO CUMPLE	OBSERVACIONES
• Eficiencia de filtración Bacteriana (BFE) EN 14683:2019+AC:2019 Anexo B			
A - Original / As received	X		
B - Tras 50 lavados / After 50 washes	X		
◆ Lavado - Mascarilla reutilizable ISO 6330:2012			
B - Tras 50 lavados / After 50 washes			VER RESULTADOS
Respirabilidad (Presión diferencial) EN 14683:2019+AC:2019 Anexo C			
A - Original / As received	X		
B - Tras 50 lavados / After 50 washes	X		

Observación: Los ensayos se realizaron según la solicitud del cliente

Código de Informe de Ensayo: AR-20-YL-008064-02-ES
Código de muestra 560-2020-00008104
Fecha de emisión 25/11/2020

LISTA DE COMPONENTES:

ID DEL COMPONENTE	NOMBRE DEL COMPONENTE	DESCRIPCIÓN DEL MATERIAL	COLOR	OBSERVACIONES
CUST 01	A - Original / As received	Textil / Fabric	Blanco / White	---
CUST 02	B - Tras 50 lavados / After 50 washes	Textil / Fabric	Blanco / White	---

Código de Informe de Ensayo: AR-20-YL-008064-02-ES
Código de muestra: 560-2020-00008104
Fecha de emisión: 25/11/2020

ENSAYOS A MASCARILLAS	CAS No.	RESULTADOS	INC.	LOQ	LÍMITES
-----------------------	---------	------------	------	-----	---------

Análisis sobre:A - Original / As received

Respirabilidad (Presión diferencial)

Fecha fin de análisis: 24/11/2020

EN 14683:2019+AC:2019 Anexo C

Presión diferencial	36.2 Pa/cm ² (± 2.0) Pa/cm ²	-	<60 Pa/cm ²	✓ Cumple
---------------------	--	---	------------------------	----------

Datos completos del ensayo incluidos en el Anexo.

• Eficiencia de filtración Bacteriana (BFE)

Fecha fin de análisis: 24/11/2020

EN 14683:2019+AC:2019 Anexo B

Eficacia de filtración bacteriana (BFE)	96.4	-	≥90 %	✓ Cumple
---	------	---	-------	----------

Datos completos del ensayo incluidos en el Anexo.
Ensayo cubierto por acreditación NABL nº TC-5492.

Análisis sobre:B - Tras 50 lavados / After 50 washes

Respirabilidad (Presión diferencial)

Fecha fin de análisis: 24/11/2020

EN 14683:2019+AC:2019 Anexo C

Presión diferencial	49.9 Pa/cm ² (± 1.5) Pa/cm ²	-	<60 Pa/cm ²	✓ Cumple
---------------------	--	---	------------------------	----------

Datos completos del ensayo incluidos en el Anexo.

◆ Lavado - Mascarilla reutilizable

Fecha fin de análisis: 24/11/2020

ISO 6330:2012

Número de ciclos de lavado	50	-
Tipo de lavadora	A1	-
Temperatura	60 °C	-
Procedimiento de lavado	6N	-
Tipo de secado	C - Secado en horizontal	-
Detergente utilizado	ECE sin fosfatos 98 + TAED + Perborato Sódico	-
Masa total seca de las probetas y el contrapeso	2 kg	-
Tipo de contrapeso utilizado	Tipo III - 100% Poliéster	-

Código de Informe de Ensayo: AR-20-YL-008064-02-ES
Código de muestra: 560-2020-00008104
Fecha de emisión: 25/11/2020

ENSAYOS A MASCARILLAS	CAS No.	RESULTADOS	INC.	LOQ	LÍMITES
-----------------------	---------	------------	------	-----	---------

- Eficiencia de filtración Bacteriana (BFE)

Fecha fin de análisis: 24/11/2020

EN 14683:2019+AC:2019 Anexo B

Eficacia de filtración bacteriana
(BFE)

95.8

-

≥90 %

✓ Cumple

Datos completos del ensayo incluidos en el Anexo.
Ensayo cubierto por acreditación NABL nº TC-5492.

Código de Informe de Ensayo:	AR-20-YL-008064-02-ES
Código de muestra	560-2020-00008104
Fecha de emisión	25/11/2020

Motivo de la nueva versión: Se modifica la fracción del ensayo tras lavados por error en registro.

Código de Informe de Ensayo: AR-20-YL-008064-02-ES
Código de muestra 560-2020-00008104
Fecha de emisión 25/11/2020

Firmado en representación de Eurofins Textile Testing Spain:



Informe validado electrónicamente por

Axel Ferrando

Responsable Laboratorio Físico/Mecánico

Nota aclaratoria

- ◆ Ensayo no amparado por la acreditación ENAC
- Test subcontratado dentro del grupo Eurofins y está acreditado
- Test subcontratado dentro del grupo Eurofins y no está acreditado
- Test subcontratado fuera del grupo Eurofins y está acreditado
- Test subcontratado fuera del grupo Eurofins y no está acreditado

N/A = No Aplica

*Eurofins Textile Testing Spain S.L.U no es responsable de la información suministrada por el cliente indicada en el apartado "Información suministrada por el cliente".

Los resultados obtenidos refieren solamente a las muestras, productos o materiales recibidos en el laboratorio, descritos en el apartado "Descripción de las muestras" y fueron ensayados en las condiciones descritas en el presente informe.

Las incertidumbres de ensayo no indicadas están a disposición del cliente, para aquellos ensayos en los que puedan ser estimadas.

La incertidumbres expandidas incluidas en el informe están basadas en la incertidumbre típica multiplicada por el factor de cobertura $k=2$, que para una distribución normal proporciona un nivel de confianza de aproximadamente el 95%

No está permitida la reproducción parcial de este documento sin la autorización expresa de Eurofins Textile Testing Spain S.L.U y cualquier impresión del mismo será considerado una copia.

Si desea hacer cualquier comentario contacte con nosotros a través de la siguiente dirección de correo electrónico **textile_spain@eurofins.com** indicando este número de informe.

FIN DEL INFORME

TEST REPORT

Test Report Issued To:

EUROFINS TEXTILE TESTING SPAIN SL

C/ GERMAN BERNACER, 9, 03203, ELCHE (ALICANTE),
ELCHE ALICANTE, SPAIN,

Test Report No: N201108002/N201108002-10

Date of Issue: 13-Nov-2020



Sample Booking/Receipt Date: 08-Nov-2020

Date of Start of Testing: 08-Nov-2020

Date of Completion of Test: 13-Nov-2020

Customer Relationship Number

60680

Sample Description :

SAMPLE DESCRIPTION-560-2020-00008104 ORIGINAL



Customer Reference No

EUAA70-00009039 ORIGINAL

Kind Attention : MELODIE RUIZ

E-Mail: MelodieRuiz@eurofins.com

Contact No: +34 966 299 638

Sample Condition : GOOD

Sample Quantity (Approx) : -

Sample Size (Approx) : -

SAMPLE NOT DRAWN BY OUR LABORATORY. THE RESULTS RELATE ONLY TO THE ITEMS TESTED

Report Issued By

Authenticity of report can be verified by mail at verification@spectrolab.in

This is a Digitally Signed Report and hence doesn't require Physical Signature.

Spectro Analytical Labs Limited, E-41, Okhla Indl. Area, Phase-II, New Delhi-110020(India)

Phone :+91-11-40522000, 41611000 || **Web**: www.spectro.in|| **Email**: care@spectro.in

ISO 9001:2015, ISO 14001:2015 & ISO 45001:2018 Certified Laboratory

Please refer to our Website <http://www.spectro.in/spectro-policies.html> for Terms & Condition

TEST REPORT FOR DETERMINATION OF:

Evaluating the Bacterial Filtration Efficiency (BFE) of Medical Face Mask Materials, Using a Biological Aerosol of Staphylococcus aureus

1. Name of Product

Fabric

2. Test Method

BS EN 14683:2019/Annex B: Medical face masks - Requirements and test methods

3. Scope

To perform bacterial filtration efficiency test on surgical mask as per procedure given BS EN 14683:2019, by using biological aerosol of Staphylococcus aureus.

4. Testing Laboratory

Name: Spectro Analytical Labs Limited
Address: E-41, Okhla Phase-2
New Delhi
Pin Code: 110020
Ph: 011-40522000

5. Date of Test

Test Date : 08.11.2020

6. Specimen Verification

6.1. Specimen Definition

The testing laboratory was not involved in the selection of the test specimen.

7. Conditioning

Each test specimens were placed for 5 h at a temperature of 21 ± 5 °C and a relative humidity of 85 ± 5 %.

8. Inoculum Size

Staphylococcus aureus ATCC 6538 (5×10^5 cfu/ml)

9. Medium Used

Tryptic soya agar

10. Temperature condition

37°C/20 to 52 hours

Rity Tiwari
Analyst Signature



Authorised Signatory

11. Observations, Requirements & Result

Sr. no.	Test Parameters	Result	Limits as per BS EN 14683:2019/ANNEX B	Test Method
1	Bacterial Filtration Efficiency of Specimen (%)	96.3	Level 1= ≥ 95 Level 2= ≥ 98 Level 3= ≥ 98	BS EN 14683:2019
		96.8		
		96.1		
		96.6		
		96.2		
	Average of BFE results (%)	96.4		
2	Area of test Specimen	Face side	Face side	BS EN 14683:2019
3	Flow rate of aerosol	28.3	28.3 L/Minute	BS EN 14683:2019
4	Mean particle size of Challenge aerosol of Staphylococcus aureus, micron	2.9	3 $\pm$ 0.3	BS EN 14683:2019
5	Average Plate count of Positive control of Staphylococcus aureus in count, per test	2322	1700-3000	BS EN 14683:2019
6	Average Plate count of Negative control	Negative	Negative	BS EN 14683:2019

Remarks: The sample showed average 96.4 % Bacterial filtration efficiency against Staphylococcus aureus ATCC 6538 according to BS EN 14683:2019 test method.

-- End of Test Report --

Rity Tiwari
Analyst Signature



Authorised Signatory

CIN:U74220DL1998PLC092698

TEST REPORT

Test Report Issued To:

EUROFINS TEXTILE TESTING SPAIN SL

C/ GERMAN BERNACER, 9, 03203, ELCHE (ALICANTE),
ELCHE ALICANTE, SPAIN,

Test Report No: N201119005/N201119005-11

Date of Issue: 24-Nov-2020



Sample Booking/Receipt Date: 19-Nov-2020

Date of Start of Testing: 19-Nov-2020

Date of Completion of Test: 24-Nov-2020

Customer Relationship Number

60680

Sample Description :

SAMPLE DESCRIPTION 560-2020-00008104



Customer Reference No

EUA70-00009039 AFTER 25 WASHES

Kind Attention : MELODIE RUIZ

E-Mail: MelodieRuiz@eurofins.com

Contact No: +34 966 299 638

Sample Condition : GOOD

Sample Quantity (Approx) : -

Sample Size (Approx) : -

SAMPLE NOT DRAWN BY OUR LABORATORY. THE RESULTS RELATE ONLY TO THE ITEMS TESTED

Report Issued By

ULR-TC549220000020887F

Authenticity of report can be verified by mail at verification@spectrolab.in

This is a Digitally Signed Report and hence doesn't require Physical Signature.

Spectro Analytical Labs Limited, E-41, Okhla Indl. Area, Phase-II, New Delhi-110020(India)

Phone :+91-11-40522000, 41611000 || **Web**: www.spectro.in|| **Email**: care@spectro.in

ISO 9001:2015, ISO 14001:2015 & ISO 45001:2018 Certified Laboratory

Please refer to our Website <http://www.spectro.in/spectro-policies.html> for Terms & Condition

TEST REPORT FOR DETERMINATION OF:

Evaluating the Bacterial Filtration Efficiency (BFE) of Medical Face Mask Materials, Using a Biological Aerosol of Staphylococcus aureus

1. Name of Product

Fabric

2. Test Method

BS EN 14683:2019/Annex B: Medical face masks - Requirements and test methods

3. Scope

To perform bacterial filtration efficiency test on surgical mask as per procedure given BS EN 14683:2019, by using biological aerosol of Staphylococcus aureus.

4. Testing Laboratory

Name: Spectro Analytical Labs Limited
Address: E-41, Okhla Phase-2
New Delhi
Pin Code: 110020
Ph: 011-40522000

5. Date of Test

Test Date : 19.11.2020

6. Specimen Verification

6.1. Specimen Definition

The testing laboratory was not involved in the selection of the test specimen.

7. Conditioning

Each test specimens were placed for 5 h at a temperature of 21 ± 5 °C and a relative humidity of 85 ± 5 %.

8. Inoculum Size

Staphylococcus aureus ATCC 6538 (5×10^5 cfu/ml)

9. Medium Used

Tryptic soya agar

10. Temperature condition

37°C /20 to 52 hours

Rity Tiwari
Analyst Signature



Testing(TC-5492)



ULR-TC549220000020887F



Authorised Signatory

11. Observations, Requirements & Result

Sr. no.	Test Parameters	Result	Limits as per BS EN 14683:2019/ANNEX B	Test Method
1	Bacterial Filtration Efficiency of Specimen (%)	96.4	Level 1= ≥ 95 Level 2= ≥ 98 Level 3= ≥ 98	BS EN 14683:2019
		95.5		
		95.7		
		95.3		
		96.1		
	Average of BFE results (%)	95.8		
2	Area of test Specimen	Face side	Face side	BS EN 14683:2019
3	Flow rate of aerosol	28.3	28.3 L/Minute	BS EN 14683:2019
4	Mean particle size of Challenge aerosol of Staphylococcus aureus, micron	2.89	3 $\pm$ 0.3	BS EN 14683:2019
5	Average Plate count of Positive control of Staphylococcus aureus in count, per test	2724	1700-3000	BS EN 14683:2019
6	Average Plate count of Negative control	Negative	Negative	BS EN 14683:2019

Remarks: The sample showed average 95.8 % Bacterial filtration efficiency against Staphylococcus aureus ATCC 6538 according to BS EN 14683:2019 test method.

-- End of Test Report --

ULR-TC549220000020887F



Rity Tiwari
Analyst Signature



Testing(TC-5492)



Authorised Signatory

METHOD FOR DETERMINATION OF BREATHABILITY (DIFFERENTIAL PRESSURE)

Test Method: EN 14683: 2019+AC: 2019 Annex C

Number of test specimens: 5


Number of test per specimen: 5

Sample area tested: Circular, diameter 2,5 cm

Tested area of the test sample: 4,9 cm²

Flow rate during testing: 8±0,25 l/min

General location of measurement areas: Representative of the overall surface.



Results original sample:

Specimen	Units (Pa)						ΔP (Pa/cm ²)
	Position 1	Position 2	Position 3	Position 4	Position 5	Mean value (Pa)	
1	167	155	162	157	167	162	33,0
2	172	174	168	160	164	168	34,2
3	186	189	175	185	188	185	37,7
4	174	188	192	187	185	185	37,8
5	191	190	187	193	183	189	38,5
						Mean Value	36,2
						Uncertainty	± 2,0

Results after 50 wash cycles at 60°C:

Specimen	Units (Pa)						ΔP (Pa/cm ²)
	Position 1	Position 2	Position 3	Position 4	Position 5	Mean value (Pa)	
1	253	248	236	241	236	243	49,6
2	259	237	228	256	249	246	50,2
3	242	236	257	230	225	238	48,6
4	239	243	247	244	248	244	49,8
5	232	246	258	257	267	252	51,4
						Mean Value	49,9
						Uncertainty	± 1,5

Observation:

For thick and rigid masks the test method may not be suitable as a proper seal cannot be maintained in the sample holder.

**SUMMARY ACCEPTANCE CRITERIA FOR REUSABLE HYGIENIC MASKS ACCORDING TO
ESPECIFICATION UNE 0065:2020**

Tests	Acceptance criteria
Bacterial filtration efficiency (BFE), (%) (Section 5.2.2 of the UNE-EN 14683: 2019 + AC: 2019 Standard)	≥ 90
Breathability (Differential pressure), (Pa/cm ²) (Section 5.2.3 of the UNE-EN 14683: 2019 + AC: 2019 Standard)	< 60

DETERMINACIÓN DE LA RESPIRABILIDAD (PRESIÓN DIFERENCIAL)

Método de ensayo: EN 14683: 2019+AC: 2019 Anexo C

Número de muestras de ensayo: 5


Número de ensayos por muestra: 5

Dimensión de la muestra de ensayo: Circular, diámetro 2,5 cm

Tamaño del área sometida a ensayo: 4,9 cm²

Caudal de aire: 8±0.25 l/min

Ubicación general de las áreas de medición: Representativos de toda la superficie



Resultados muestra original:

Muestra de ensayo	Unidades (Pa)						ΔP (Pa/cm ²)
	Posición 1	Posición 2	Posición 3	Posición 4	Posición 5	Valor medio	
1	167	155	162	157	167	162	33,0
2	172	174	168	160	164	168	34,2
3	186	189	175	185	188	185	37,7
4	174	188	192	187	185	185	37,8
5	191	190	187	193	183	189	38,5
						Media	36,2
						Incertidumbre	± 2,0

Resultados después de 50 ciclos de lavado a 60°C:

Muestra de ensayo	Unidades (Pa)						ΔP (Pa/cm ²)
	Posición 1	Posición 2	Posición 3	Posición 4	Posición 5	Valor medio	
1	253	248	236	241	236	243	49,6
2	259	237	228	256	249	246	50,2
3	242	236	257	230	225	238	48,6
4	239	243	247	244	248	244	49,8
5	232	246	258	257	267	252	51,4
						Media	49,9
						Incertidumbre	± 1,5

Observación:

Para mascarillas gruesas y rígidas, el método de ensayo puede no ser adecuado dado que no se pueden mantener un sello apropiado en el soporte de la muestra.

**RESUMEN CRITERIOS DE ACEPTACIÓN PARA MASCARILLAS HIGIÉNICAS REUTILIZABLES
SEGÚN UNE 0065:2020**

Ensayos	Criterio de aceptación
Eficacia de filtración. (%) (<i>“ensayo BFE”</i>) Ensayada a través de ensayo de eficacia de filtración bacteriana (BFE), (%) (apartado 5.2.2 de la Norma UNE-EN 14683:2019+AC:2019)	≥ 90
Respirabilidad (Presión diferencial), (Pa/cm ²) (apartado 5.2.3 de la Norma UNE-EN 14683:2019+AC:2019)	< 60